

We, ABB Limited

Of, Oldends Lane Stonehouse Gloucestershire England GL10 3TA

Declare under our own responsibility that the product: **Gas Monitor Type 6553** to which this declaration relates is in conformity with the following standards:

Directive 2006/95/EC (LVD)

EN61010-1:2010

Directive 2004/108/EC (EMC)

Harmonised Standard Used: EN 61326-1:2006

Directive 94/9/EC (ATEX)

Harmonised Standard Used: EN 60079-0:2011+A11:2013 EN 60079-11:2011

Provisions of the Directive fulfilled by the Equipment. Equipment Group:-



II (1)G [Ex ia Ga] IIC (-20°C≤Ta ≤+ 40°C)

Name, notified body number and address of the Notified Body:-Baseefa Ltd, Number 1180, of:- Rockhead Business Park, Staden Lane, Buxton, Derbyshire SK17 9RZ

Number of the type examination certificate:-BAS01ATEX7043/3 - last supplement issued 15th January 2016

When installed and operated according to the supplied instructions.

12/10/16

Date of Issue

Signature Y.Saulick Certification Engineer